

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Uppatti

Lab No 12468

[Details About Robust Analysis](#)

Year 2023

[Detail About Monthly Summary](#)

Month February

[Detail about SDI](#)

Constituent Group Chemistry III

[All Method Result](#)[Print](#)

Result Entered Date : 22/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	153.37	8.21	12.59	152 mg/dL	-0.11	4.26
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	11	29.98	5.87	1.76	37 mg/dL	3.99	1.06
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	37	3.35	7.70	0.26	3.4 mg/dL	0.19	0.08
4	CHOLESTROL III	CHOD-PAP III	Biosystems	33	113.39	6.52	7.39	118 mg/dL	0.62	2.57

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Home](#)

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Uppatti

Lab No 12468

Year 2023

Month March

Constituent Group Chemistry III

[Details About Robust Analysis](#)[Detail About Monthly Summary](#)[Detail about SDI](#)[All Method Result](#)[Print](#)

Result Entered Date : 20/03/2023

Sl.No	Analyte	Method / Principle Name	Anal.yser Name	No of Participant's	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	92.67	8.46	7.84	94 mg/dL	0.17	2.61
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	12	55.01	12.13	6.67	56 mg/dL	0.15	3.85
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.37	9.41	0.13	1.4 mg/dL	0.23	0.04
4	CHOLESTROL III	CHOD-PAP III	Biosystems	37	112.13	5.96	6.68	109 mg/dL	-0.47	2.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Home		

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Uppatti	
Lab No	12468	Details About Robust Analysis
Year	2023	Detail About Monthly Summary
Month	April	Detail about SDI
Constituent Group	Chemistry III	

[All Method Result](#)
[Print](#)

Result Entered Date : 20/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67	136 mg/dL	2.00	2.93
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	12	41.88	8.81	3.69	40 mg/dL	-0.51	2.13
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15	2.2 mg/dL	0.14	0.05
4	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03	123 mg/dL	1.70	2.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Home		

VIEW LAB MONTHLY SUMMARY

Handwritten notes:
 286
 801
 SDI
 All

Lab Name Govt. PHC, Uppatti
 Lab No 12468
 Year 2023
 Month May
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

All Method Result

Print

Result Entered Date : 20/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85	176 mg/dL	-0.39	3.43
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	11	25.79	11.91	3.07	24 mg/dL	-0.58	1.85
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16	1.79 mg/dL	0.06	0.05
4	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47	124 mg/dL	1.00	2.49

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Home		

Handwritten signature and stamp:
 Govt. P.H.C. Uppatti
 Pambalur T.K.
 District, Erode

CROSS CHECK SLIDES PARTICULARS

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Uppatti
 Lab No: 12468
 Year: 2023
 Month: June
 Constituent Group: Chemistry III

[All Method Result](#) [Print](#)

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Result Entered Date : 16/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	280.51	4.07	11.42	269.77 mg/dL	-0.94	3.81
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	12	31.96	9.97	3.19	30.41 mg/dL	-0.49	1.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	39	2.67	6.50	0.17	2.66 mg/dL	-0.06	0.06
4	CHOLESTROL III	CHOD-PAP III	Biosystems	37	113.79	4.23	4.81	123.37 mg/dL	1.99	1.58

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Home		

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Uppatti

Lab No 12468

[Details About Robust Analysis](#)

Year 2023

[Detail About Monthly Summary](#)

Month July

[Detail about SDI](#)

Constituent Group Chemistry III

[All Method Result](#)[Print](#)

Result Entered Date : 20/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.10	105.93 mg/dL	-0.45	1.70
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	11	50.80	9.44	4.79	56.4 mg/dL	1.17	2.89
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34	6.12 mg/dL	0.81	0.12
4	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.04	107.86 mg/dL	-1.45	1.39

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Home		