

CALIBRATION CERTIFICATE

Date: 27.09.2023

Calibration Name & Address		CERTIFICATE NO: AWM09C23	
The Medical Officer Government Primary Health Centre Onnupuram.			
Details of Device Under Calibration (DUC)			
Description	: SEMI AUTO ANALYZER	Make	: Robonik
Range	:	Model	: Pritest Touch
Least Count	:	Sr.No	: ATCD1001220RBK
DUC Condition	: Satisfactory	Location	: LAB
Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25°C	Relative humidity	49%RH
Sample Calibration Date	06/03/2023	Calibration Done at	ON SITE

RESULTS

S.No	Specification	Measured Values in Ω	Allowable limit in Ω	Uncertainty in Ω	Remarks
1	Earth Bond Resistance	0.260	<2 Ω	0.02	✓PASS/FAIL
		Measured Values in M Ω	Allowable limit in M Ω	Uncertainty in M Ω	Remarks
2	Insulation Resistance	44.90	>2 M Ω	5.26	✓PASS/FAIL
		Measured Values in $\mu\Omega$	Allowable limit in $\mu\Omega$	Uncertainty in $\mu\Omega$	Remarks
3	Earth Leakage (NC)	214	<5000 $\mu\Omega$ for B, BF,CF	19.35	✓PASS/FAIL
4	Earth Leakage (SFC)	248	<1000 $\mu\Omega$ for B, BF,CF		✓PASS/FAIL
		Measured values in $\mu\Omega$	Allowable limit in $\mu\Omega$	Uncertainty in $\mu\Omega$	Remarks
5	Enclosure Leakage (NC)	5	<1000 $\mu\Omega$ for B, BF,CF	3.57	✓PASS/FAIL
6	Enclosure Leakage (SFC)	272	<500 $\mu\Omega$ for B, BF,CF	19.76	✓PASS/FAIL

Remarks

- 1.This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory
- 2.The user Should be determine the suitability of the instrument for its intended use
- 3.The Recalibration interval should be determine on the User requirement.
- 4.The results Stated in this Certificate relate only to the item Calibrated
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00
- 6.Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By



Authorized Signatory

