

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Sembalur  
 Lab No: 7189  
 Month: February  
 Year: 2023  
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 26/02/2023

Date of Report Published : 08/03/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1868	155.11	6.56	10.18	152.1 mg/dL	-0.30	0.47
2	CREATININE III	Jaffe's Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1872	3.40	9.00	0.31	3.1 mg/dL	-0.98	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1788	112.79	5.23	5.90	104.1 mg/dL	-1.47	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*S. Srinivas*

MEDICAL OFFICER  
 GOVT. PRIMARY HEALTH CENTRE  
 SEMBANLUR - 630 313  
 SIVAGANGAI DISTRICT

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. Block PHC, Sembanur	<b>Details About Robust Analysis</b>
<b>Lab No</b>	7169	<b>Detail About Monthly Summary</b>
<b>Month</b>	March	<b>Detail about SDI</b>
<b>Year</b>	2023	
<b>Constituent Group</b>	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	104.1 mg/dL	1.85	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	115.1 mg/dL	0.72	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*C. Smithy*  
 MEDICAL OFFICER  
 GOVT. PRIMARY HEALTH CENTRE,  
 SEMBANUR - 630 313  
 SIVAGANGAI DISTRICT

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Sembanur	<a href="#">Details About Robust Analysis</a>
Lab No	7169	<a href="#">Detail About Monthly Summary</a>
Month	April	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 25/04/2023

Date of Report Published : 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Your Value	SDI	SI
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	122.1 mg/dL	0.27	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	1.8 mg/dL	-2.66	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.51	5.19	5.79	119.2 mg/dL	-0.24	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Analysis</a>	<a href="#">Modify Methods</a>	<a href="#">Duplicate Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Weekly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Logout</a>

*G. Srinivas*

MEDICAL OFFICER  
 GOVT. PRIMARY HEALTH CENTRE  
 SEMBANUR - 630 313  
 CHANGANGAL DISTRICT

023, 20 59



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. Block PHC, Sembanur  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 7169  
Date of Result Entered: 20/05/2023  
Date of Report Published: 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1034	180.88	5.19	9.39	190 mg/dL	0.97	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.5 mg/dL	-1.97	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	6.08	5.83	116.1 mg/dL	0.26	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. Block PHC, Sembanur  
The Block Medical Officer, Sembanur, Sivagangal District  
Sembanur  
TAMILNADU 630313

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0418-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
MEDICAL OFFICER  
GOVT PRIMARY HEALTH CENTRE  
SEMBANUR - 630 313  
SIVAGANGAL DISTRICT

7/2023, 21 06



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name: Govt. Block PHC, Sembanur  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 7169  
Date of Result Entered: 20/06/2023  
Date of Report Published: 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1923	256.63	5.17	14.82	255.1 mg/dL	-0.10	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.3 mg/dL	-1.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.48	5.07	105.1 mg/dL	-1.70	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. Block PHC, Sembanur  
The Block Medical Officer, Sembanur, Sivagangal District  
Sembanur  
TAMILNADU 630313

Coordinator Contact Details:  
Email: cllnqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*Opki*  
GOVT. BLOCK PHC, SEMBANUR  
SIVAGANGAL DISTRICT  
TAMILNADU - 630313

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Semtanur

Details About Robust Analysis

Lab No 7159

Detail About Monthly Summary

Month July

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

S.No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-PDD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	117.1 mg/dL	1.28	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.73	0.00	0.35	5.3 mg/dL	-1.51	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.68	127.2 mg/dL	3.05	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution, Warning Signal.
Beyond ±3.0	Unacceptable performance, Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Summary
Yearly Summary	Change Password	Home	Settings

*C. Chandra*

MEDICAL OFFICER  
GOVT. PRIMARY HEALTH CENTRE  
SEMANTUR, P.O. 313  
SIVAMANGAL DISTRICT

Department of Clinical Biochemistry  
Christian Medical College, Vellore - 632004



**CMC EXTERNAL QUALITY ASSURANCE SCHEME**  
(Under the aegis of ACBI)

**Certificate of Participation**

*This is to certify that*  
**GOVT. BLOCK PRIMARY HEALTH CENTER**

.....  
**SEMBANUR**  
.....

*Lab No. 7169 has participated*  
*in the Chemistry III Program of the External Quality Assurance Scheme*  
*during the year 2022..*

*Pamela Christudoss*

**Dr. Pamela Christudoss, M.Sc., Ph.D.,**

EQAS Coordinator

Professor of Clinical Biochemistry