

VIEW LAB MONTHLY SUMMARY

Lab Name **Govt. PHC, Koodacheri**
 Lab No **12446**
 Month **March**
 Year **2023**
 Constituent Group **Chemistry III**

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	94.32 mg/dL	0.42	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.5 mg/dL	1.40	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	108.2 mg/dL	-0.47	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

MEDICAL OFFICER,
 Govt. Primary Health Centre,
 KOODACHERY - 637 200,
 P. Velur(Tk), Namakkal.

Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)