

B MONTHLY SUMMARY

Name: Govt. Block PHC, Athani  
 No: 7368  
 Month: June  
 Year: 2023  
 Constituent Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	272 mg/dL	-0.99	0.69
CREATININE III	Jeffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.2 mg/dL	-2.16	0.01
CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	110 mg/dL	-0.74	0.24
SDI Range Interpretation									
Within -1.00 to +1.00	Excellent.								
Within ±1.01 to ±2.00	Good.								
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.								
Outside ±3.0	Unacceptable performance. Action Signal.								

Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Summary	Change Password	Home	Sign-out