

CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



Govt. Block PHC, Athani

Lab No

7368

Chemistry III

Date of Result Entered :

23/05/2023

Lyophilized human serum based

Date of Report Published :

05/06/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
DOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	184 mg/dL	0.33	0.44
ININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
STROL III	C. OD-PAP III	Rob vnik (Semi automatic / Fully Automatic)	17,8	114.60	5.08	5.43	107 mg/dL	-1.30	0.27
Interpretation									
SDI Range									
+1.00			Excellent.						
±2.00			Good.						
±2.99			Accept with caution. Warning Signal.						
			Unacceptable performance. Action Signal.						

Dr. Athani
MEDICAL OFFICER, CLINICAL LABORATORY, ERODE DISTRICT.

Pamela Christodoss

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Contact Details:

சென்னை மருத்துவ கல்லூரி
சென்னை மருத்துவ கல்லூரி
சென்னை மருத்துவ கல்லூரி
சென்னை மருத்துவ கல்லூரி
சென்னை மருத்துவ கல்லூரி