



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. PHC, Paramathi
 Constituent Group Chemistry III
 PT item Lyophilized human serum based

Lab No 12447
 Date of Result Entered : 19/06/2023
 Date of Report Published : 04/07/2023

S.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	300 mg/dL	0.90	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	107 mg/dL	-1.33	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Paramathi
 CLINICAL LABORATORY
 Paramathi
 TAMILNADU637207

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

V. D. J. 15/7/23
 மருத்துவ அலுவலர்,
 அரசினர் ஆரம்ப கல்வியியல் நிகழ்வு
 மாமகி, (காமகம்) மாமகி, (காமகம்)

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential

ps://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult_new.aspx?resDetId=266&noSamples=1&month=6&year=2023&groupId=9&yearGroup=110