



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name Govt. PHC, Paramathi Lab No 12447
Constituent Group Chemistry III Date of Result Entered : 20/03/2023
PT item Lyophilized human serum based Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	D
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	91 mg/dL	-0.07	0.31
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	175	55.20	11.70	6.46	53.2 mg/dL	-0.31	0.98
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.39 mg/dL	0.50	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	103 mg/dL	-1.37	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Paramathi
CLINICAL LABORATORY
Paramathi
TAMIL NADU 637207

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

[Handwritten Signature]
மருத்துவ ஆய்வகம்,
அரசினர் ஆரம்ப சுகாதார நிலையம்,
பரமத்தி, (தாயக்கல் மாவட்டம்).