



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023



PC-102A

Lab Name: Govt. Block PHC, Athani  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 7368  
Date of Result Entered: 18/03/2023  
Date of Report Published: 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	99 mg/dL	1.10	0.31	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01	
3	C.OLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.9	5.23	5.80	110 mg/dL	-0.16	0.28	
SDI Range				Interpretation							
Within -1.00 to +1.00				Excellent.							
Within ±1.01 to ±2.00				Good.							
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.							
Beyond ±3.0				Unacceptable performance. Action Signal.							

LAB ADDRESS :  
Govt. Block PHC, Athani  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, ERODE DISTRICT.  
Athani  
TAMILNADU638502

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Paula Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*Handwritten signature and stamp*  
மாதிரி சேகரிக்கப்பட்டது  
04/04/2023  
0416-2283102