

## CALIBRATION CERTIFICATE

Date: 25.09.2023

Calibration Name & Address	Certificate No	AWMS25C23	
Government Primary Health Centre Mukkur			
Details of Device Under Calibration (DUC)			
Description : Micro Pipette	Make: Thermascientific		
Range : 10 - 100, 100 - 1000	Model : F3		
Least Count : ±2	Sr.No : 4640040 , 4640060.		
DUC Condition :	Location : Lab		
Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25°C	Relative humidity	49%RH
Sample Calibration date	06.03.23	Calibration Done at	ON SITE

### CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS			DEVIATION	EXPANDED UNCERTAINTY(±)2ul
	ul					
1	100 - 1000ul	1000ul	1000ul	1000ul	Ok	Ok
2	10 - 100ul	100 ul	100ul	100ul	Ok	Ok

### Remarks

- 1.This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory
- 2.The user Should be determine the suitability of the instrument for its intended use
- 3.The Recalibration interval should be determine on the User requirement.
- 4.The results Stated in this Certificate relate only to the item Calibrated
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00
- 6.Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By

(N)

Authorized Signatory



*M. A. An.*