



<b>CALIBRATION REPORT</b>	
Report No : HBMS/CAL/23-24	Calibration Date: 25.09.2023
	Calibration valid: 24.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, URBAN PRIMARY HEALTH CENTRE, TIRUCHENCODE, NAMAKKAL.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMOSCIENTIFIC
ii Model	FINNPIPETTE F3
iii Value	100µl-1000µl
iv Serial No	RW02482

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023





**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

**For Heart Beat Medical Systems,**

**(Authorized signatory)**

