



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC INTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. PHC, Poigai Lab No: 12798
 Constituent Group: Chemistry III Date of Result Entered: 18/08/2023
 PT Item: Lympholized human serum based Date of Report Published: 01/09/2023

Sl. No.	Analyte	Method / Analytical System	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.09	157 mg/dL	-2.06	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	9.12	1.97 mg/dL	1.76	0.01
3	CHOLESTROL III	GHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.89	4.07	4.64	169 mg/dL	-1.04	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Poigai
 Poigai
 TAMILNADU

Coordinator Contact Details:
 Email: clinq@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

PPA
8/10/23
 MEDICAL OFFICER
 Govt. Primary Health Centre
 POIGAI - 627 856.