

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Mookanamalaipatti									
Lab No	12503	Back to Search Page								
Month	July									
Year	2023	Print								
Constituent Group	Chemistry III									
Date of Result Entered : 19/07/2023										
Date of Report Published : 07/08/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.10			1.70
			Transasia / Erba	24	109.83	6.83	7.50			3.06
			Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	120 mg/dL	1.76	0.28
			Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95			1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34			0.12
			Transasia / Erba	18	5.79	3.00	0.17			0.08
			Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.9 mg/dL	0.20	0.02
			Any Analyser (Automation / Semi Automation)	66	5.89	4.60	0.27			0.07
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.04			1.39
			Transasia / Erba	19	112.64	4.38	4.93			2.26
			Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	118 mg/dL	1.07	0.22
			Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88			0.80
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out