

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Mookanamalaipatti									
Lab No	12503	Back to Search Page								
Month	June									
Year	2023	Print								
Constituent Group	Chemistry III									
Date of Result Entered : 20/06/2023										
Date of Report Published : 04/07/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	280.51	4.07	11.42			3.81
			Transasia / Erba	22	283.77	5.80	16.44			7.01
			Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	288 mg/dL	0.09	0.69
			Any Analyser (Automation / Semi Automation)	65	287.55	4.50	12.95			3.21
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	39	2.67	6.50	0.17			0.06
			Transasia / Erba	23	2.50	9.65	0.24			0.10
			Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
			Any Analyser (Automation / Semi Automation)	67	2.58	6.39	0.16			0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	113.79	4.23	4.81			1.58
			Transasia / Erba	22	114.61	5.43	6.22			2.65
			Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	112 mg/dL	-0.34	0.24
			Any Analyser (Automation / Semi Automation)	149	114.61	4.05	4.64			0.76
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out