

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Mookanamalaipatti									
Lab No	12503	Back to Search Page								
Month	April									
Year	2023	<input type="button" value="Print"/>								
Constituent Group	Chemistry III									
Date of Result Entered : 20/04/2023										
Date of Report Published : 06/05/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67			2.93
			Transasia / Erba	19	118.42	8.48	10.04			4.61
			Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	115 mg/dL	-0.62	0.34
			Any Analyser (Automation / Semi Automation)	74	118.33	5.77	6.83			1.59
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15			0.05
			Transasia / Erba	19	2.14	6.03	0.13			0.06
			Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	118 mg/dL	643.56	0.01
			Any Analyser (Automation / Semi Automation)	65	2.14	7.02	0.15			0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03			2.07
			Transasia / Erba	20	111.15	4.25	4.72			2.11
			Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	1.9 mg/dL	-18.95	0.28
			Any Analyser (Automation / Semi Automation)	143	111.22	4.31	4.79			0.80
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out