


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230708A/N-116	Page	: 1 of 1
CALIBRATED ON.	: 10 / 07 / 2023	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 09 / 07 / 2024	C.P. No.	: ACS/CP/10
SRF No. & DATE	: 230708A/N Dt. 08.07.2023	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342723000000749F Dt. 13.07.2023	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. M R Road, Besides IDBI Bank, Bylane S F Road, Siliguri 734005		
ITEM CALIBRATED	: Centrifuge	Make	: Remi
		Range	: (0 to 6000) RPM
		L. Count	:
		ID No.	: RCL/BIO/Cent-01
		Location	: Clinical Biochemistry
		S. No.	: ZIFN-27243
ENVIRONMENT CONDITION:	Temp. { 25°C ± 4°C } R.H. {30% to 75%}		
STANDARD INSTRUMENTS USED :			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.
DUE DATE			
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1
			12/09/2023
RESULTS	: Mechanical Discipline		
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.16	- 1.16	
2000	2002.20	- 2.20	
4000	4005.38	- 5.38	
6000	6006.34	- 6.34	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.60 rpm of reading at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by *Rej*

....End of Calibration Certificate....

S.K. Sharma
 Auth. Signatory
 S.K. Sharma
 CEO/TM