


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230708A/N-121	Page	: 1 of 1	
CALIBRATED ON.	: 10 / 07 / 2023	Ref. Std.	: CG-11	
CALIBRATION DUE.	: 09 / 07 / 2024	C.P. No.	: ACS/CP/07	
SRF No. & DATE	: 230708A/N Dt. 08.07.2023	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342723000000754F Dt. 13.07.2023	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. M R Road, Besides IDBI Bank, Bylane S F Road, Siliguri 734005			
ITEM CALIBRATED	: Digital Temperature Indicator			
	Make : RoHS	Range	: (-50 to 110) °C	
	L. Count :	ID No.	: RCL/HEM/D.Temp-04	
	Location : Haematology	S. No.	: N.A	
ENVIRONMENT CONDITION:	Temp. { 25°C + 4°C} R.H. {30% to 75%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	10/11/2023
RESULTS	: ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
2.0	2	0.0		
4.1	4	0.1		
6.2	6	0.2		
8.3	8	0.3		
(average of readings)				

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is $\pm 1.35^\circ\text{C}$ at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

