


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO. : 230708A/N-102 Page : 1 of 1

CALIBRATED ON. : 08 / 07 / 2023 Ref. Std. : ISO:8655:6  
CALIBRATION DUE. : 07 / 07 / 2024 C.P. No. : ACS/CP/08

SRF No. & DATE : 230708A/N Dt. 08.07.2023 Phy. Inspection : O.K.

ULR No. & ISSUE DATE : CC342723000000735F Dt. 13.07.2023 Cal. At : Site / Lab

CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd.  
M R Road, Besides IDBI Bank, Bylane S F Road,  
Siliguri 734005.

ITEM CALIBRATED : Micro Pipette  
Make : Thermo Scientific Range : (10 to 100) µl  
L. Count : ID No. : RCL/HEM/PIP-02  
Location : Haematology S. No. : TW08034

ENVIRONMENT CONDITION: Temp. { 25°C ± 4°C } R.H. {30% to 75%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/398/JAN/23	05/01/2024

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.11	(-) 0.11
100		50	50.18	(-) 0.18
		100	100.37	(-) 0.37

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.1 µl at coverage factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

