



CALIBRATION / TESTING REPORT

Table with customer details, certificate no, equipment name, manufacture, model, calibration date, and due date.

- 1. General maintenance required
2. Any Physical damage observed
3. Spare Replacement (if required)

Table with test equipment, manufacture, model, and serial number.

TEST REPORT

Temperature 25°C
Humidity RH 55%

Table with parameters, units, set value, measured values, tolerance, and status.


*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By :

Handwritten signature

Authorized Sign With Seal



*Electrical safety test result attached

CALIBRATION / TESTING REPORT			
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE SATHARASANKOTTAI, SIVAGANGAI.		CERTIFICATE NO	TICS/2023/MM/SG/01
		EQUIPMENT NAME	CENTRIFUGE
		MANUFACTURE	LABOTECH
		MODEL	-
CALIBRATION DATE	27/09/2023	ASSET NO	SKLAB-001
CALIBRATION DUE DATE	26/09/2024	INSTALLED DEPARTMENT	LAB

TEST REPORT –Electrical Safety

Test Standard: IEC 62353

Parameter	Units	Value	Limits	Status
Main Voltage L1-L2	V	232.9	240±10	PASS
Main Voltage L1-G(PE)	V	232.3	240±10	PASS
Main Voltage L2-G(PE)	V	0.5	<5	PASS
Load Current	A	0.405	As per unit	PASS
Protective Earth Resistance	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0.21	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- <50μA*	PASS

* Note: Defibrillator paddles 100 μA, per IEC 60601-2-4

The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

