

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Vellore  
 Lab No 12204  
 Month May  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)[Detail About Monthly Summary](#)[Detail about SDI](#)[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)


Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	174 mg/dL	-0.73	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	112 mg/dL	-0.45	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Change Password](#)[Home](#)[Sign-out](#)


MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 VELLORE