

ALLWIN MEDICAL SYSTEMS

No. 2/4, 2nd Street, 1st Floor, Jayalakshmi Nagar, Kattupakkam, Chennai - 600 056, Tamilnadu, India,

Cell: 9443663366

Email: allwinmedicalsystems@gmail.com

CALIBRATION CERTIFICATE

Date: 29.09.2023

Calibration Name & Address

CERTIFICATE NO: AWM13C23

The Medical Officer

Government Primary Health Centre,

Nedumbirai.

Details of Device Under Calibration (DUC)

Description

: SEMI AUTO ANALYZER

Make: Genuine Bio System

Range

SEIVII AOTO AIVAETZER

Model: GB235+

Least Count
DUC Condition

: Satisfactory

Sr.No: ES1031812P9097

Location : LAB

Environmental Conditions & Calibration Procedure Details

64%RH

Environmental Details
Sample Calibration Date

Temperature: 25°C 06/03/2023

Relative humidity
Calibration Done at

ON SITE

RESULTS

S.No	Specification	Measured Values in Ω	Allowable limit in Ω	Uncertainty in Ω	Remarks
1	Earth Bond Resistance	0.288	<2Ω	0.02	PASS/FAIL
		Measured Values in MΩ	Allowable limit in $M\Omega$	Uncertainty in $M\Omega$	Remarks
2	Insulation Resistance	44.63	>2 MΩ	5.62	PASS/FAIL
	_	Measured Values in μΩ	Allowable limit in µA	Uncertainty in µA	Remarks
3	Earth Leakage (NC)	217	<5000 µA for B, BF,CF	18.78	PASSIFAIL
4	Earth Leakage (SFC)	435	<1000µAfor B, BF,CF		PASS/FAIL
		Measured values in μA	Allowable limit in µA	Uncertainty in µA	Remarks
5	Enclosure Leakage (NC)	6	<1000µAfor B, BF,CF	3.04	PASS/FAIL
6	Enclosure Leakage (SFC)	278	<500µAfor B, BF,CF	21.04	PASS/FAIL

Remarks

- 1. This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory
- 2. The user Should be determine the suitability of the instrument for its intended use
- 3. The Recalibration interval should be determine on the User requirement.
- 4. The results Stated in this Certificate relate only to the item Calibrated
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00
- 6.Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By

EDITAL

thorized Signatory

CHENNAI- 56