



CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE SATHARASANKOTTAI, SIVAGANGAL.		CERTIFICATE NO		TICS/2023/MM/SG/06	
		EQUIPMENT NAME		MICROPIPETTE	
		MANUFACTURE		--	
		MODEL		10-100 µL	
CALIBRATION DATE	27/09/2023	ASSET NO		SKLAB-006	
CALIBRATION DUE DATE	26/09/2024	INSTALLED DEPARTMENT		LAB	
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT		ID. NO		CERTIFICATE NUMBER	
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01		TSC/23-24/5424-1	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	50	50.02	± 1	PASS
		100	999.16	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : Authorized Sign With Seal					
					

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE SATHARASANKOTTAI, SIVAGANGAI.		CERTIFICATE NO		TICS/2023/MM/SG/05	
		EQUIPMENT NAME		MICROPIPETTE	
		MANUFACTURE		--	
		MODEL		100-1000µL	
CALIBRATION DATE	27/09/2023	ASSET NO		SKLAB-005	
CALIBRATION DUE DATE	26/09/2024	INSTALLED DEPARTMENT		LAB	
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT		ID. NO		CERTIFICATE NUMBER	
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01		TSC/23-24/5424-1	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	99.82	± 1	PASS
		1000	999.91	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : Authorized Sign With Seal					
