

		CALIE	BRAT	10	N CER	TIF	ICA	TE_		
CERTIFICATE NO:	SBS/CL/2				MECHANICAL			Page	No:1 of 1	
Issue Date					29-09-2023					
SRF No & Date					SRF/23/00464-0001 & 28-09-2023					
Receipt Date					28-09-2023					
Calibration Date					28-09-2023					
Calibration Due					27-09-2024					
Customer Name & A	ddress									
GOVERNMENT PRIM	IARY HEALTH	CENTRE,								
SEMBANUR-630305,	SIVAGANGAI [	DISTRICT.								
		Deta	ils of Dev	vice	Under Calibr	ration	(DUC)			
Description : CENTRIFUGE		JGE		Make	e & Model		;	MICROSIL &	MK-10	
Range :	ange : 3400 RPM			Sr. N	lo		Ė	239/385		
Resolution : 100 RPM		12	Ident	tification No		2	NA			
DUC Condition :	Satisfacto	ry		Loca	tion		:	LAB		
	Envi	onmental Co	onditions	& S	tandard Ope	erating	Proce	dure Details		
Environmental Details Temperature			e: 25.5°C F		Relative Humidity		52%Rh			
Calibration Procedure No SBS/CP/ML			04 Calibrat		Calibration d	done at		ONSITE		
			Refere	nce :	Standards D	etails				
S.No Description	S.No Description			Make/ SI No:		Certificate No				Validity
1 Digital Tachometer			LINE SE	LINE SEIKI / 175-0034V			JRPM-CCTR-A&S-2023-0013			09-06-2024

## **CALIBRATION RESULTS**

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY		
	RPM	RPM	RPM	%		
1	1000	999.0	1.0	5.8		
2	2000	1999.5	0.5	5.8		
3	3000	2999.5	0.5	5.8		

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By
(Calibration Engineer)
P.PRASANNA



Authorised Signatory

(Guarity Manager) (Chief Executive)

(C.SIVABALAN)

## **SUNSHINE BIOMEDICAL SOLUTIONS**