



RAJA MEDICAL SYSTEMS

No.15, V.O.C STREET, COIMBATORE - 641 011. MOBILE : 96263 13796.

Calibration Protocol & Report for Medical Equipment

Equipment : SEMI AUTO ANALYSER
 MAKE : ROBONIK
 MODEL : PRIETEST TOUCH

PROTECT CLASS /TYPE : Risk Class I
 S. No. : AT2800418RBK
 Location : GPHC, VELLODE

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	230 volts 0 to 5 volts No Zero ohms 10mA NA	Stabilizer/UPS With in Ranges Well
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Front Panet/Top lid, buckets, motor assembly & power cord (c) Risk of mechanical damages, if any	No OK No	Checked OK
3	Performance Test (a) On /Off switch (b) switch & functional tests (c) Running test (d) Abnormalities (e) Display checkup (f) Keypad/indicators & connector	Normal OK OK No Well Well	Ok Perfect Perfect
4	Calibration Test (a) End Point Test (b) Display checkup (c) Filter selection (d) Record OD (e) Calculation (f) Two Point Test (g) Kinetic Test (h) Result	Done Ok OK OK Well Perfect	Perfect
Formulated By: Einstein Research Laboratory		Issued By:	
Endorsed By : Research team head		RAJA MEDICAL SYSTEMS No.15, V.O.C STREET, COIMBATORE - 641 011.	
Reviewed On: 10.02.23 Next Review : 09.02.24		User Department:	