

## VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name **Govt. PHC, Rapposal**Lab No **12502**[Back to Search Page](#)Month **April**Year **2023**[Print](#)Constituent Group **Chemistry III**Date of Result Entered : **19/04/2023**Date of Report Published : **06/05/2023**

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67			2.93
			Transasia / Erba	19	118.42	8.48	10.04			4.61
			Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	128 mg/dL	1.19	0.34
			Any Analyser (Automation / Semi Automation )	74	118.33	5.77	6.83			1.59
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15			0.05
			Transasia / Erba	19	2.14	6.03	0.13			0.06
			Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	1.7 mg/dL	-2.63	0.01
			Any Analyser (Automation / Semi Automation )	65	2.14	7.02	0.15			0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03			2.07
			Transasia / Erba	20	111.15	4.25	4.72			2.11
			Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	113 mg/dL	0.24	0.28
			Any Analyser (Automation / Semi Automation )	143	111.22	4.31	4.79			0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

