



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL CHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE ACADEMY
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. PHC, Sethumadi
Constituent Group: Chemistry III
PT Item: Lymphoidized human serum based

Lab No: 0209
Date of Result Entered: 19/08/2023
Date of Report Published: 01-09/2023

2023.10.04.14:43

Sl. No.	Analyte	Method / Principle Name	Analyzer Name	No. of Substances	DV	Participants		Your Value	SCN	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	8.68	178 mg/dL	0.27	0.42
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.20	15.20	4.30	21 mg/dL	-1.70	0.66
3	CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.78	6.77	0.12	1.6 mg/dL	-1.34	0.01
4	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97	106 mg/dL	-1.68	0.81

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±3.00	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Sethumadi
CLINICAL LABORATORY, COIMBATORE
Sethumadi
TAMILNADU-642131

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Rudh
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
சேத்துமடை