



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. UPHC, Neelikonampalayam

Constituent Group Chemistry III

Lab No

10758

PT item

Lyophilized human serum based

Date of Result Entered :

19/08/2023

Date of Report Published :

01/09/2023

| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 9.00 | 163 mg/dL | -1.40 | 0.42 |
|-----------------------|----------------|--|--|------|--------|------|------|-----------|-------|------|
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.8 mg/dL | 0.34 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1764 | 113.80 | 4.07 | 4.64 | 119 mg/dL | 1.12 | 0.22 |
| SDI Range | | Interpretation | | | | | | | | |
| Within -1.00 to +1.00 | | Excellent | | | | | | | | |
| Within ±1.01 to ±2.00 | | Good. | | | | | | | | |
| Within ±2.01 to ±2.99 | | Accept with caution. Warning Signal. | | | | | | | | |
| Beyond ±3.0 | | Unacceptable performance. Action Signal. | | | | | | | | |

LAB ADDRESS :

Govt. UPHC, Neelikonampalayam
DR. MURUGESAN, MEDICAL OFFICER, NEELIKONAMPALAYAM, NEAR PERIYA VEEDHI, SAMUDHAYA KOODAM
Coimbatore
TAMILNADU641003

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential