



SILICON TECHNOLOGIES

2 & 3, Devi Illam, Priya Nagar,
Telungupalayam Pirivu,
Coimbatore - 641 039, TamilNadu.

GSTIN : 33APLPA1828A1ZM

T. ARUN PRASAD, M.Tech., MBA., (Ph.D)
Managing Director
Consultant Biomedical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : SEMI AUTO ANALYSER Protect Class/Type : Risk class II/CF

Make : ROBONIK S.No : AT2150211RBK

Model : PRIETEST TOUCH Location : PUNNAM - GPHC

S.no	Measurement Parameters	Status	Remarks
I	Electrical Test (a). Safety Standards (range: 210-240 V) (b). Equipotential ground & Earthing (c). Risk of shock/fire Hazards, if any (d). Chassis & Insulation Resistance (e). Earth & Chassis Leakage Current (f). Battery condition & charging status	235volts 0 to 5volts No Zero ohms 10mA -	Stabilizer/UPS With in ranges well
II	Physical Condition Test (a). Cabinet & Visible damages, if any (b). Switches, display, optics chamber, Power cord accessories (c). Risk of mechanical damages, if any	No Ok No	Checked ok
III	Performance Test (a). biochemistry test (b). photo detector test (c). Optics performance (d). Abnormalities (e). Display checkup (f). Keypad/indicators & connectors	Normal Ok /wave length No Ok Well	Fine Clarity perfect nm
IV	Calibration Test (a). Input power regulation done (b). SMPS power supply various outputs -/+3V,5V,12V calibration done (c). Power supply to source Cal -I (d). Source lamp Sensor Cal-II (range: 1250-1300ohms) (e). Detector sensitivity (range: 370-940nm)	Tuned calibrated done 6VDC Done 1295 ohms Normal	tuned in range well
Formulated By : Einstein Research Laboratory		Issued By:	User Department:
Endorsed By : Research team head		SILICON TECHNOLOGIES Biomedical Instrumentation Company COIL. JATORE - 641 039 Cell: 98652 85754, 81246 75020.	
Reviewed On : 28.06.23 Next Review : 28.12.23			

Email : st.biomedical@gmail.com

☎ 98652 85754, 81246 75020