



SILICON TECHNOLOGIES

2 & 3, Devi Illam, Priya Nagar,
Telungupalayam Pirivu,
Coimbatore - 641 039, TamilNadu.

T. ARUN PRASAD, M.Tech., MBA., (Ph.D)
Managing Director
Consultant Biomedical Engineer

GSTIN : 33APLPA1828A1ZM

Calibration Protocol & Report for Medical Equipment

Equipment: CENTRIFUGE Protect Class/Type : Risk class I

Make : LABOTECH S.No : 20111278

Model : LAB152 Location : PUNNAM - GPHC

S.no	Measurement Parameters	Status	Remarks
I	Electrical Test (a). Safety Standards (range: 210-240 V) (b). Equipotential ground & Earthing (c). Risk of shock/fire Hazards, if any (d). Chassis & Insulation Resistance (e). Earth & Chassis Leakage Current (f). Battery condition & charging status	230volts 0 to 5volts No Zero ohms 10mA NA	Stabilizer/UPS With in ranges well
II	Physical Condition Test (a). Cabinet & Visible damages, if any (b). Front panel/top lid, buckets, motor assembly & Power cord (c). Risk of mechanical damages, if any	No Ok No	 Checked ok
III	Performance Test (a). On/off switch (b). Speed regulator switch & functional tests (c). Running test (d). Abnormalities (e). Display checkup (f). Keypad/indicators & connectors	Normal ok accurate Ok No Well Well	ok perfect perfect
IV	Calibration Test (a). Vibration test (b). Speed regulation gradual hike/wobbling tests (c). Running test with load (d). Display checkup (f). timer test	Done ok —	perfect

Formulated By : Einstein Research Laboratory		Issued By:	User Department:
Endorsed By : Research team head	<i>Tap</i>	SILICON TECHNOLOGIES Biomedical Instrumentation Company COIMBATORE - 641 039 Cell: 98652 85754, 81246 75020.	
Reviewed On : 28.06.23			
Next Review : 28.12.23			