



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. PHC, Sethiathope

Lab No: 9212

Constituent Group: Chemistry III

Date of Result Entered: 20/06/2023

PT Item: Lyophilized human serum based

Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	268 mg/dL	-1.26	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.4 mg/dL	-0.99	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	108 mg/dL	-1.13	0.24
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Sethiathope
THE MEDICAL OFFICER, SETHIATHOPE PHC CHIDAMBARAM TALUK, CUDDALORE-
Sethiathope
TAMILNADU608702

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Handwritten notes:
 2/2
 Nov:- 4.19
 (Good)
 Glucose: ✓
 Creatinine: ✓

Signature:
 Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore