



| <b>CALIBRATION REPORT</b>                           |  |
|---|--|
| Report No : HBMS/CAL/23-24                          | Calibration Date : 12.09.2023  |
|   | Calibration valid : 11.03.2024   |
| <b>1.1.CUSTOMER DETAILS:</b>                        |  |
| Name and Address of the organization                | <b>M/s, GOVERNMENT PRIMARY HEALTH CENTRE ,<br/>UKKIRANKOTTAI, THIRUNELVELI</b> |
| <b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b> |  |
| EQUIPMENT NAME                                      | CENTRIFUGE   |
| <b>i</b> Manufactured by                            | MICROSIL   |
| <b>ii</b> Model                                     | MK-10  |
| <b>iii</b> Serial No                                | 118-286  |

**CALIBRATION WORKSHEET**

**LOCATION: LABORATORY**

**Tabulation For – Centrifuge**

| S.no | Unit under test IN Minutes | Observed speed tachometer in RPM | Deviation measured $\pm$ RPM |
|------|----------------------------|----------------------------------|------------------------------|
| 1    | 1000                       | 1000                             | 0                            |
| 2    | 1200                       | 1201                             | 1                            |
| 3    | 1500                       | 1501                             | 1                            |
| 4    | 2000                       | 2003                             | 3                            |

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB



99400 28355, 9751206106

88833 15662



No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com