

**CALIBRATION REPORT**

Report No: HBMS/CAL/23-24

Calibration Date: 23.09.2023

Calibration Valid: 22.03.2024

**I.1. CUSTOMER DETAILS:**

Name and Address of the organization

M/S CORPORATION HOSPITAL  
UPHC, NELLUKONAMPALAYAM, COIMBATORE**I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):**

EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	MERILETT
ii	Model	-
iii	Value	100µl-1000µl
iv	Serial No	339537

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.06	1000.01
2	100.03	500.05	1000.03
3	100.05	500.02	1000.07
4	100.09	500.10	1000.08
5	100.00	500.02	1000.09
6	100.08	500.07	1000.02

For Heart Beat Medical Systems,

  
 (Authorized signatory)

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsistem@gmail.com