

### CALIBRATION REPORT

Report No: HBMS/CAL/23-24	Calibration Date: 23/09/2023
	Calibration Valid: 22/03/2024
<b>I.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/S CORPORATION HOSPITAL, UPHC, NEELIKONAMPALAYAM, COIMBATORE
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i) Manufactured by	THERMO SCIENTIFIC
ii) Model	FINEFITTE - F3
iii) Value	10 µl -100 µl
iv) Serial No	RW01642


**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report**

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.11
2	10.04	50.09	100.08
3	10.05	50.02	100.10
4	10.02	50.08	100.08
5	10.00	50.02	100.09
6	10.08	50.07	100.11

**For Heart Beat Medical Systems,**

  
(Authorized signatory)