



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name	Govt. Block PHC, Kedar	Lab No	7272
Constituent Group	Chemistry III	Date of Result Entered :	18/06/2023
PT item	Lyophilized human serum based	Date of Report Published :	04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	265.3 mg/dL	-1.44	0.69
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	50	33.99	14.82	5.04	27 mg/dL	-1.39	1.42
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.53 mg/dL	-0.23	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	107 mg/dL	-1.33	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Kedar

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.

Kedar

TAMILNADU605402

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****