



## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Chittalandur

[Details About Robust Analysis](#)

Lab No 12455

[Detail About Monthly Summary](#)

Month August

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

| Sl.No                 | Analyte                        | Method / Principle Name                  | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value | SDI   | U    |
|-----------------------|--------------------------------|--|--|--------------------|--------|--------------|------|------------|-------|------|
|                       |                                |  |  |                    |        | CV           | SD   |            |       |      |
| 1                     | <a href="#">GLUCOSE III</a>    | GOD-POD III                              | Robonik ( Semi automatic / Fully Automatic ) | 1841               | 175.58 | 5.13         | 9.00 | 176 mg/dL  | 0.05  | 0.42 |
| 2                     | <a href="#">CREATININE III</a> | Jaffes Kinetic - Alkaline Picrate III    | Robonik ( Semi automatic / Fully Automatic ) | 1852               | 1.76   | 6.77         | 0.12 | 1.8 mg/dL  | 0.34  | 0.01 |
| 3                     | <a href="#">CHOLESTROL III</a> | CHOD-PAP III                             | Robonik ( Semi automatic / Fully Automatic ) | 1764               | 113.80 | 4.07         | 4.64 | 113 mg/dL  | -0.17 | 0.22 |
| SDI Range             |                                | Interpretation                           |  |                    |        |              |      |            |       |      |
| Within -1.00 to +1.00 |                                | Excellent.                               |  |                    |        |              |      |            |       |      |
| Within ±1.01 to ±2.00 |                                | Good.                                    |  |                    |        |              |      |            |       |      |
| Within ±2.01 to ±2.99 |                                | Accept with caution. Warning Signal.     |  |                    |        |              |      |            |       |      |
| Beyond ±3.0           |                                | Unacceptable performance. Action Signal. |  |                    |        |              |      |            |       |      |

| Links          |                 |                    |                          |
|----------------|-----------------|--------------------|--------------------------|
| Modify Address | Modify Methods  | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home               | Sign-out                 |

Site Maintained by  
 **Swiftware Solutions**

Best Viewed e