



**VIEW LAB MONTHLY SUMMARY**

**Lab Name** Govt. PHC, Chittalandur [Details About Robust Analysis](#)  
**Lab No** 12455 [Detail About Monthly Summary](#)  
**Month** July [Detail about SDI](#)  
**Year** 2023  
**Constituent Group** Chemistry III

[Click on the analyte to view Graphical Data](#)
[All Analyser Result](#)
[Print](#)

**Date of Result Entered :** 16/07/2023  
**Date of Report Published :** 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	114 mg/dL	0.77	0.28
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.5 mg/dL	1.91	0.02
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	110 mg/dL	-0.64	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

*Site Maintained by*  
**Swiftware Solutions**
Best Viewed e