



**VIEW LAB MONTHLY SUMMARY**

<b>Lab Name</b>	<b>Govt. PHC, Chittalandur</b>	<a href="#">Details About Robust Analysis</a>
<b>Lab No</b>	<b>12455</b>	<a href="#">Detail About Monthly Summary</a>
<b>Month</b>	<b>April</b>	<a href="#">Detail about SDI</a>
<b>Year</b>	<b>2023</b>	
<b>Constituent Group</b>	<b>Chemistry III</b>	

[Click on the analyte to view Graphical Data](#)

**All Analyser Result**
**Print**

Date of Result Entered : 17/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	124.4 mg/dL	0.69	0.34
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.28 mg/dL	0.69	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	110 mg/dL	-0.28	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Site Maintained by

Best Viewed e