



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - MARCH 2023

PC-1024

Lab Name Govt. PHC, Avaloor

Lab No

12218

Constituent Group Chemistry III

Date of Result Entered :

20/03/2023

PT item Lyophilized human serum based

Date of Report Published :

04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD				
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	94 mg/dL	0.37	0.31	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.1 mg/dL	-1.90	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	111 mg/dL	0.01	0.28	
<b>SDI Range</b>											
				<b>Interpretation</b>							
Within -1.00 to +1.00				Excellent.							
Within ±1.01 to ±2.00				Good.							
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.							
Beyond ±3.0				Unacceptable performance. Action Signal.							

**LAB ADDRESS :**  
 Govt. PHC, Avaloor  
 THE MEDICAL OFFICER, KANCHEEPURAM  
 Kancheepuram  
 TAMILNADU631605

**Coordinator Contact Details:**  
 Email:cilingqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christodoss*  
 Dr. Pamela Christodoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

*M. V. S. R.*  
 MEDICAL OFFICER,  
 GOVT. PHC.,  
 AVALUR - 631 605.  
 KANCHIPURAM DISTRICT

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential