



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name: Govt. PHC, Melaramanathi

Lab No: 12559

Constituent Group: Chemistry III

Date of Result Entered: 20/03/2023

PT Item: Lyophilized human serum based

Date of Report Published: 04/04/2023

Sl. No.	Test Name	Result	Unit	Reference Range	Interpretation					
1	GLUCOSE III	GOD-POD III	74	85 mg/dL	-1.05	1.66				
2	CREATININE III	Jeffes Kinetic - Alkaline Picrate III	70	1.32	7.52	0.10	1.1	mg/dL	-2.22	0.02
3	CHOLESTROL III	CHOD-PAP III	148	111.55	4.29	4.78	163	mg/dL	-1.79	0.79
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

LAB ADDRESS :

Govt. PHC, Melaramanathi

Melaramanathi

TAMILNADU

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential