

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Vadamadurai
 Lab No 4590
 Month April
 Year 2023
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

All Analyser Result

Print

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl. No.	Analyte	Principle Name	Analyte Method	No of Participants	CV	Participants		Your Value	SDI	
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	101 mg/dL	-2.56	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.2 mg/dL	0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	109 mg/dL	-0.45	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Example: [http://www.viewMonthlySummary/viewResult2_new.aspx?resDetId=264&icSamples=1&month=4&year=2023&groupId=9&yearGroup=110](#)

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Vadamadurai
 Lab No 4590
 Month May
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	196 mg/dL	1.61	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.9 mg/dL	1.06	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	109 mg/dL	-0.96	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

.inks

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Vadamadurai **Details About Robust Analysis**
Lab No 4590 **Detail About Monthly Summary**
Month June **Detail about SDI**
Year 2023
Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		You		
						CV	SD	Val		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	265 mg/dL	-1.46	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	104 mg/dL	-1.92	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly
Yearly Summary	Change Password	Home	Sign out

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Vadanadurai
 Lab No: 2590
 Month: July
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analysis to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 15/07/2023

Date of Report Published : 07/08/2023

Sr	ANALYSIS	REAGENT	ANALYSER	CONC	UNIT	SDI	SDI	SDI	SDI	SDI	SDI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	102	mg/dL	1.21	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.9	mg/dL	0.20	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	118	mg/dL	1.07	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within = 1.01 to = 2.00	Good
Within = 2.01 to = 2.99	Accept with caution. Warning Signal
Beyond = 3.0	Unacceptable performance. Action Signal.

Links

பஞ்சவர்ண ஆய்வகம்
 சேலம் சிவில் ஸ்டாப்டு ரோடில்
 வி.பஞ்சவர்ண - 624 808

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Vadamadurai
 Lab No 4590
 Month August
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	187 mg/dL	1.27	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	2 mg/dL	2.02	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	117 mg/dL	0.69	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

//home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=268&noSamples=1&month=8&year=2023&groupId=9&yearGroup=110