

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Aruvankadu	Details About Robust Analysis
Lab No	10563	Detail About Monthly Summary
Months	June	Detail about SDI
Year	2023	
Consultant Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	313 mg/dL	1.78	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	192	32.69	12.96	4.24	36 mg/dL	0.78	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	126 mg/dL	2.42	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

K. Anand

MEDICAL OFFICER,

GOVT. PRIMARY HEALTH CENTRE,

ARUVANKADU