




# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name: Govt. PHC, Pillikkal Palayam  
 Constituent Group: Chemistry III  
 PT Item: Lyophilized human serum based  
 Lab No: 12432  
 Date of Result Entered: 20/04/2023  
 Date of Report Published: 06/05/2023

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	119.5 mg/dL	0.01	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.02 mg/dL	-0.80	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	116.7 mg/dL	0.88	0.28
<b>SDI Range</b>										
				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**  
 Govt. PHC, Pillikkal Palayam  
 CLINICAL LABORATORY  
 Pillikkal Palayam  
 TAMILNADU637213

**Coordinator Contact Details:**  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential