

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Puzhikkal Palayam
 Lab No 12432
 Month March
 Year 2023
 Constituent Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

S/No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonk (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	92.12 mg/dL	0.09	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonk (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.44 mg/dL	0.91	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonk (Semi automatic / Fully Automatic)	1761	110.93	6.23	5.80	112.16 mg/dL	0.21	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within +1.01 to +2.00				Good.						
Within +2.01 to +2.99				Accept with caution. Warning Signal.						
Beyond +3.0				Unacceptable performance. Action Signal.						

Links

- Modify Address [Modify Methods](#) [Submit Lab Results](#) [View Lab Monthly Summary](#)
- Yearly Summary [Change Password](#) [Home](#) [Sign-out](#)