

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Aruvankadu

Lab No: 10563

Month: June

Year: 2023

Department Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonk (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	313 mg/dL	1.78	0.69
2	UREA III	Ureaso UV / GLDH III	Robonk (Semi automatic / Fully Automatic)	192	32.69	12.96	4.24	36 mg/dL	0.78	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonk (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonk (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	126 mg/dL	2.42	0.24
SDI Range										
				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

K. Anand

MEDICAL OFFICER.

GOVT. PRIMARY HEALTH CENTRE

ARUVANKADU