

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12554

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
THERIKUVELMUTHUKULATHUR(TK),
RAMANATHAPURAM DISTRICT-623711

SRF No.	SRF/23/00558-0001
SRF Date	26-09-2023
Date of Receipt	25-09-2023
Date of Calibration	26-09-2023
Due Date for Calibration	25-09-2024
Issue Date	30-09-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW13644	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.7	Humidity (%RH)	50	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023


Z Factor: 1.00319

Result of Calibration in µl


Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	100.51	100.52	100.53	100.52	100.51	100.50	0.50	0.02	0.47
		100.49	100.48	100.49	100.49	100.48				
2	500	500.40	500.42	500.43	500.44	500.46	500.43	0.43	0.02	0.47
		500.45	500.44	500.43	500.42	500.41				
3	1000	1001.24	1001.26	1001.23	1001.22	1001.22	1001.22	1.22	0.03	0.47
		1001.18	1001.17	1001.20	1001.24	1001.26				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor $k=2$ (95).
- Calibration Liquid Used: Distilled or Deionized water conforming Grade 1 as specified in ISO 3696.

Calibrated by: 
(Calibration Engineer)
M.P.A.C.27



Authorised by: 
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.