



CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	
Calibration Date: 28.08.2023	
Calibration Valid: 27.02.2024	
I.1. CUSTOMER DETAILS:	
Name and Address of the organization	
M/s, GOVT. PRIMARY HEALTH CENTRE, FILIKKALPALAYAM, NAMAKKAL.	
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	
PIPETTE	
I	Manufactured by
THERMOSCIENTIFIC	
II	Model
FINNEPIPETTE F3	
III	Value
1000µl	
IV	Serial No
GPHC/PPM/MP/002	

Calibration standards used :


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel I (Volume)
	1000µl
1	1000.05
2	1000.06
3	1000.01
4	999.97
5	1000.01
6	1000.02

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com