



CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 28.08.2023
	Calibration Valid: 27.02.2024
I.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVT. PRIMARY HEALTH CENTRE, PILIKKALPALAYAM, NAMAKKAL.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I Manufactured by	THERMOSCIENTIFIC
II Model	FINNEPIPETTE F3
III Value	100µl
IV Serial No	GPHC/PPM/MP/001


Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	100µl
1	100.05
2	100.06
3	100.01
4	99.97
5	100.01
6	100.02

For Heart Beat Medical Systems,



(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com