



## CALIBRATION CERTIFICATE

Page No: 1 of 1

Certificate No: SBS/CL/22/07974

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,  RAMESHWARAM MAIN ROAD, NARIKUDI-626607.	SRF No.	SRF/22/00182-0027 &
	SRF Date	18-08-2022
	Date of Receipt	18-08-2022
	Date of Calibration	20-08-2022
	Due Date for Calibration	19-08-2023
	Issue Date	20-08-2022

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	QW13939	Operating Range	
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LAB

Environmental Condition

Calibration Method Used

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	15112918	TVCSPL21/12/1587-01	A&D & GH-252	National Standards	10-12-2022

Z Factor: 1.00319

Result of Calibration in µl

SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.93	99.92	99.93	99.92	99.93	99.87	-0.13	0.06	0.47
		99.81	99.82	99.81	99.82	99.81				
2	500	499.94	499.95	499.94	499.95	499.94	499.93	-0.08	0.02	0.47
		499.91	499.90	499.91	499.90	499.91				
3	1000	999.84	999.85	999.84	999.85	999.84	999.85	-0.15	0.01	0.47
		999.87	999.86	999.84	999.86	999.87				

Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)  
**S.LAKSHMANAN MURUGAN**



Authorised by:

(Quality Manager/Chief Executive)  
**D.VETRI SELVI**

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First floor , Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.