

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Aarapur	Details About Robust Analysis
Lab No	10575	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

SL No.	Analyte	Method / Principle Name	Analyser Name	No. of Participants	CV	Participants			SDI	SI
						CV	SD	Mean Value		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	92 mg/dL	0.08	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	110 mg/dL	-0.16	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Muddy Methods	Submit Lab Results	View Lab Monthly Summary
Monthly Summary	Change Password	Home	Logout