

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12055		MEDICAL DEVICES		Page No:1 of 1
Issue Date	29-09-2023			
SRF No & Date	SRF/23/00467-0002 & 28-09-2023			
Receipt Date	28-09-2023			
Calibration Date	28-09-2023			
Calibration Due	27-09-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, APPANUR-623703.RAMANATHAPURAM DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH	
Range	MULTI	Sr. No	ATCD3050321RBK	
Resolution	MULTI	Identification No	NA	
DUC Condition	SATISFACTORY	Location	LAB	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature 25.6 °C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024


ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 79	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B BF CF	Measured values in µA 209	Uncertainty in % (±) 7.3
3	Enclosure Leakage <500µAfor B BF CF	Measured values in µA 229	Uncertainty in % (±) 7.2


REMARKS

- 1 This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2 The user should determine the suitability of the instrument for its intended use.
- 3 The recalibration interval should be determined on the user requirement
- 4 The results stated in this certificate relate only to the item calibrated.
- 5 The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6 Equipment used for Calibration were calibrated & traceable to National & International Standards

Calibrated By

(Calibration Engineer)
P.PRASANNA



Authorised Signatory

 Technical Manager
C.SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.